

Venture Medical ReQuip, Inc. 6008 Bonacker Drive Tampa, FL 33610 800-627-3215 | 813-740-8640 Fax: 800-466-1251 sales@venturemedical.com

Biolight M8000A Color Patient Monitor -New



The Biolight M8000A Patient Monitor features a 10.4" color LCD display with user friendly menus and a navigator knob for easy configurations.

The M8000A is a perfect solution for recovery, PACU, procedure rooms and anywhere a compact multi-parameter monitor is needed.

Standard configuration includes: ECG (3 or 5 lead), SpO2, NIBP, Resp, PR, 2-Temp (skin or esophageal), Co2 Ready, Printer, and VGA Output.

The Biolight M8000A includes: ECG cable and leadwires (3 lead), NIBP cuff, NIBP hose, SpO2 finger sensor, lithium battery, recorder paper, and user manual.

Optional rolling stands and wall mounts are also available.

Manufacturer Warranty: 2 Years, 90 Days for Accessories

Technical Specifications:

- Display: 10.4" Color LCD
- Dimension: 12.5" (W) x 10.4" (H) x 6" (D)
 Weight: ≤ 9.9 lbs. (with battery)
- Weight: ≤ 9.9 lbs. (with battery) • Power Supply: 120~240V; AC: 50~ 60 Hz
- Alarm Indication: Visual and Audible Alarms

Trend

- Store & Review: 120 Hours, Trend Data and Trend Maps
- Parameter Options: SpO2, NIBP, PR, Resp, Co2, Temp 1-2, IBP 1-2, ST
- Storage Cycle Intervals: 1- 5min, 10min, 15min, 20min, 25min, 30min

Parameter Specifications: ECG

- Leads: I, II, III, aVR, aVL, aVF
- Mode: 3-leads, 5-leads
- Gain: 2.5mm/mV, 5mm/mV, 10mm/mV, 20mm/mV, 40mm/mV, auto
- Scanning Speed: 12.5mm/s, 25mm/s, 50mm/s
- Heart Rate Range:
 - Adult: 10~300bpm
 - Child & Neonate: 10~350bpm
- Accuracy: ±1bpm or ±1% whichever is greater
- <u>ST</u>
- Measurement Range: -2.0mV 2.0mV
- Resolution: 0.01mV
- <u>NIBP</u>
 - Operation Modes: Automatic Oscillometry
 - Range: Adult: 10 ~270mmHg Child: 10 ~235mmHg
 - Neonate: 10 ~135mmHg
 - Measurement: Systolic Pressure/Diastolic Pressure/Mean Pressure
 - Unit: mmHg/Kpa selectable
 - Pulse rate range: 40 ~240bpm
- RESP
 - Method: Impedance variation between RA-LL (R-F)
 - Measuring Range: 0~150rpm
 - Accuracy: ±2rpm
 - Gain: x1, x2, x4
 - Sweep speed 6.25mm/s, 12.5mm/s, 25mm/s
- Biolight SpO2 (Pulse Oximetry)
 - Range: 0 ~100%
 - Accuracy: ±2% (70% ~100%); Unspecified (0 ~69%)
 - Range & Accuracy of Pulse Rate: 25 ~250bpm ±1bpm
 - or ±1% whichever is greater

```
TEMP
```

- Channel: 2 channels
 Range: 32~122°F, 0~50°C
- Unit: Celsius, Fahrenheit
- Standard Configuration:
 - ECG, RESP, NIBP, SpO2, PR, Temp, VGA Output, Co2 Ready, Rechargeable Lithium Battery, Printer
- IBP (option)
 - Channel: 2
 - Measurement Range: -50 ~ +300mmHg
 - Accuracy: ±2mmHg or 2%, whichever is greater
- EtCo2 (option, Sidestream)
 - Range 0 ~19.7% (0 ~150mmHg)
 - Unit: %,mmHg, kPa
 - Respiration Rate Range 2~150bpm

VENTUREMEDICAL.COM